



## SMC chiller system and power supply installation

# SMC Chiller unit (HRS024-A-20 model) set up



The Chiller system is running properly without water leakage and maintain the setting temp correctly .

# Laser DC power supply unit set up



The power on without fault as green indicator .  
The DC 48V power supply output 48V to laser head normally .

